

Diaphragm Seals**PIP #: DS/PI-49A**Applicable to
All Diaphragm Seals**ASHCROFT DIAPHRAGM SEALS
ON CHLORINE APPLICATIONS**

Halocarbon fill fluid must be used on all chlorine applications. If the process media includes chlorine in any form, Halocarbon is required. Examples are:

- Chlorine (Dry)
- Chlorine (Moist)
- Chlorine Vapors
- Chlorine Dioxide
- Chlorine Trifluoride
- Etc.

If a liquid filled gauge is attached to the seal, the fill fluid must again be Halocarbon.

Consult the corrosion guide for material recommendations. If chemicals are not listed, contact Ashcroft Customer Service at 203-378-8281.