

Wahl DST 600 Competition Comparison Chart			
Feature	Wahl DST 600	Dart Reference Thermometer (Anderson)	DTG Digital Thermometer (Anderson)
Meter Accuracy	+ 0.30°F over one year	+ 0.50°F over one year	+ 0.50°F over one year
Ambient Operating Limits	-40 to 158°F (-40 to +70°C)	40 to 120°F (5 to 49°C)	40 to 140°F (5 to 60°C)
Battery Life	Minimum 24 mths	Line Powered	Minimum 6 months
Display Size	1.0" Display	0.5" Display	0.5" Display
Display Update	Factory Set at 1 second, adjustable in .25 second intervals from .25 to 10 seconds	None stated	3 seconds
NIST Certified	Yes	Yes	Yes
Quick Disconnect Remote cable connections for easy calibration	YES – can remove meter & probe to lab for calibration of temperature system	NO	NO – cannot be easily removed after installa- tion, in field requires two people, two sensors, and 3 temperature baths to calibrate as a system.
Calibration Pts	ONE – ice point	None Stated	Two up to Five
Patent Pending Continuous Self Checking Technology	YES – continuous confirmation of meter accuracy	NO – unable to confirm accuracy of meter	NO – unable to confirm accuracy of meter
Housing NEMA 4-X	YES	NO	YES
Adjustable Angle Mounting System	YES	NO	NO
Tamper Resistant	YES – can only be adjusted using factory supplied software	YES – can only be adjusted using factory supplied software	NO – have to use mechanical device not sold with DTG



(800) 421-2853 • FAX (828) 658-0728 • www.palmerwahl.com